




# Powder Free Nitrile

## EXAMINATION GLOVES



## High Quality Exam Glove with pH5.5 Technology!

Our **Lighthouse Series Powder Free Nitrile** gloves work with your hands allowing a full range of motion, while helping to lessen the effect of hand fatigue. Our **Lighthouse W211** gloves with pH5.5 Technology, are the professionals' choice for a wide variety of applications. Strong, durable and value priced!

- **Breakthrough pH5.5 Technology!** Made to match healthy skin's natural pH5.5 levels to ensure the **MOST** comfortable fit yet!
- **Unique formulation reduces excess perspiration**
- **Textured for maximum grip in both wet and dry conditions**
- **Nitrile formulation alleviates the potential for adverse reactions associated with natural rubber latex proteins**
- **Approved for Medical Use**
- **Powder Free**
- **Ambidextrous**

Item Number	Size
W211-XS . . . . .	X-SMALL
W211-S . . . . .	SMALL
W211-M . . . . .	MEDIUM
W211-L . . . . .	LARGE
W211-XL . . . . .	X-LARGE

100 gloves/box,  
10 boxes/case

<sup>1</sup> **Caution:** Components used in making these gloves may cause allergic reactions in some users. Follow your institution's policies for use.



A family of hand protection and safety products under one roof.



## POWDER FREE NITRILE EXAM GLOVES SPECIFICATIONS

### MATERIAL

Nitrile

### OUTER SURFACE

Textured

### DONNING POWDER

Free from glove powder

### COMPONENTS

100% Acrylonitrile-butadiene (nitrile), a synthetic co-polymer

### SHAPE

Ambidextrous, straight fingers, thumb and fingers in one plane. Beaded cuff.

### SIZES

X-Small (XS), Small (S), Medium (M), Large (L), Extra Large (XL)

### PACKAGING

1000 gloves per case, by weight, 10 boxes per case 100 gloves per dispenser box, by weight

### COLOR

Blue

### PIN HOLES AND VISUAL DEFECT

Assurance Action Sampling Inspection by destructive and visual test procedures

Each glove is 100% air tested to ensure superior quality

### GOOD MANUFACTURING PRACTICE

The gloves are manufactured in compliance with the Current Good Manufacturing Practice (GMP)

Requirements in the United States of America as appropriate for patient examination gloves

(Federal Register, Part 820)

### ISO 9000

Quality System Certified to International Standards Organization 9002. Model for Quality Assurance in Production, Installation and Servicing. Accreditation Number: QSC-4674

### COUNTRY OF ORIGIN

China

### QUALITY CHARACTERISTICS

#### Physical Dimensions

<b>Width</b>	X-Small	73 mm	+/- 4 mm
	Small	79 mm	+/- 4 mm
	Medium	89 mm	+/- 4 mm
	Large	104 mm	+/- 4 mm
	X-Large	114 mm	+/- 4 mm
<b>(inches / mm)</b>			
<b>Length</b>		min. 9.5 / 235	+/- 10 mm
	<b>(mils / mm)</b>		
<b>Thickness</b>	Cuff	3.75 / 0.1	+/- 0.04 mm
	Palm	4.9 / 0.13	+/- 0.04 mm
	Fingertips	6.0 / 0.15	+/- 0.04 mm

©2013 ATLANTIC SAFETY PRODUCTS, INC. ALL RIGHTS RESERVED. 2013.180